

DISSOLVED OXYGEN TRANSMITTER

Model : TR-DOT1A4

ISO-9001, CE, IEC1010



勝特力材料 886-3-5753170
勝特力电子(上海) 86-21-34970699
勝特力电子(深圳) 86-755-83298787
[Http://www.100y.com.tw](http://www.100y.com.tw)



Lutron

LUTRON ELECTRONIC

The Art of Measurement

DISSOLVED OXYGEN TRANSMITTER

Model : TR-DOT1A4

| | |
|------------------------------|--|
| Measure Range | <p><i>Range 1 :</i> O² CAL.</p> <p><i>Range 2 :</i> DO : 0 -20 mg/L</p> |
| Output | <p>4 - 20 mA</p> <p><i>Range 1 (CAL) :</i> 20.9 % O² = 12 mA</p> <p><i>Range 2 (DO) :</i> 0 mg/L = 4 mA 10 mg/L = 12 mA 20 mg/L = 20 mA</p> |
| External Adjustment | ZERO VR SPAN VR |
| Power Supply | 90 - 260 ACV 50 Hz/60 Hz |
| Operation Temperature | 0 - 50 °C 32 - 122 °F |
| Operation Humidity | Less than 80 %RH |
| Accessory Included | Dissolved Oxygen Probe Probe-filling electrolyte, OXEL-03 Spare diaphragms, OXDP-02 |
| Optional Accessories | Probe-filling electrolyte, OXEL-03 Spare diaphragms, OXDP-02 |
| Mounting | Din rail or Wall |
| Case | ABS plastic |
| Others | Separate probe with plug |

* Appearance and specifications listed in this brochure are subject to change without notice.

勝特力材料 886-3-5753170
 勝特力电子(上海) 86-21-34970699
 勝特力电子(深圳) 86-755-83298787
[Http://www.100y.com.tw](http://www.100y.com.tw)